

Calibration of Signal Conditioners

Remote Charge Converter IEPE100

Item	Description
IEPE100-AK	IEPE100 – Traceable calibration at fixed frequency points (0.2 Hz – 50 kHz) • Incoming inspection to verify the calibration capability • Calibration at reference frequency of 160 Hz for each gain range • Calibration at 14 additional octave-spaced frequencies from 0.2 Hz to 50 kHz Calibration method according to DKD-R 3-2: 04/2019 Documentation: accredited calibration certificate according to DIN EN ISO/IEC 17025:2018

